



## REVIEW ARTICLE

# Nasal high flow therapy for neonates: Current evidence and future directions

Kate A Hodgson, Peter G Davis and Louise S Owen

Newborn Research Centre, Royal Women's Hospital, Melbourne, Victoria, Australia

**Abstract:** Nasal high flow (nHF) therapy is a commonly used method of providing non-invasive respiratory support for neonates. It has several potential mechanisms of action: continuous distending pressure, nasopharyngeal dead space washout, provision of heated and humidified gases and reduction of work of breathing. nHF is used in a number of clinical scenarios for preterm and term infants, including primary respiratory and post-extubation support. In recent years, large trials have generated evidence pertinent to these indications. Novel applications for nHF in neonates warrant further research: during endotracheal intubation, for initial delivery room stabilisation of preterm infants and in conjunction with minimally invasive surfactant therapy.

**Key words:** nasal high flow; neonatology; respiratory.

The 1960s saw the advent of endotracheal ventilation support for preterm infants. During this era, mortality rates were high, and complications such as pneumothorax, pulmonary interstitial emphysema and bronchopulmonary dysplasia (BPD) were common. Post-mortem examinations identified hyaline membrane disease,<sup>1</sup> and subsequently, surfactant deficiency was identified as the underlying cause.<sup>2</sup> The availability of exogenous surfactant therapy in the 1990s revolutionised neonatal intensive care in Australia, reducing morbidity and mortality.<sup>1,3</sup> Despite this, the incidence of BPD remained high, and research turned to non-invasive methods of respiratory support for these infants. Initially, observational studies suggested a reduction in the incidence of BPD when nasal continuous positive airway pressure (CPAP) was used instead of mechanical ventilation.<sup>4,5</sup> Randomised studies<sup>6–8</sup> subsequently showed that CPAP could be used from birth, even in extremely preterm infants. Infants managed with

CPAP are less likely to receive surfactant, have fewer days of endotracheal ventilation and may have a lower risk of BPD.<sup>9,10</sup> CPAP has now become the mainstay of non-invasive support for preterm infants, although it has some disadvantages: nasal trauma is common,<sup>11</sup> the interface is bulky, and highly trained nursing staff are required to manage CPAP. In recent years, an alternative, nasal high flow (nHF) therapy, has greatly increased in popularity.<sup>12</sup> nHF uses smaller bi-nasal prongs and delivers heated, humidified gas at flows of >1 L/min, typically 2–8 L/min.<sup>12</sup> Its widespread use in neonates occurred before evidence of its safety and efficacy was available. Recently, several large trials have evaluated nHF in various clinical settings. With respect to the clinical applications of nHF, this review will present the available meta-analysis and randomised controlled trial evidence for nHF. Due to the relative paucity of evidence regarding mechanisms of action of nHF, these sections have included studies conducted with a variety of research methodologies.

## Key Points

- 1 The use of nasal high flow (nHF) for neonates has increased greatly in recent years.
- 2 nHF may be an effective alternative to continuous positive airway pressure for post-extubation support of preterm infants >28 weeks' gestation.
- 3 There is conflicting evidence regarding the use of nHF for primary respiratory support of preterm infants and practice will depend upon many factors.

**Correspondence:** Dr Kate A Hodgson, Newborn Research Centre, Royal Women's Hospital, 20 Flemington Road, Parkville, Vic. 3052, Australia. Fax: +61 38345 3489; email: kate.hodgson@thewomens.org.au

Conflict of interest: None declared.

Accepted for publication 16 December 2018.

## Mechanisms of Action

### Continuous distending pressure

During CPAP therapy, continuous distending pressure is thought to improve lung mechanics and prevent atelectasis.<sup>13</sup> In contrast to CPAP, where the clinician sets the applied distending pressure, nHF requires *flow* rather than *pressure* to be selected. Although nHF does generate some distending pressure, this varies with gas flow and leak. Despite initial concerns about excessive pressure generation and risk of pneumothorax,<sup>14,15</sup> most *in vivo* studies have shown that pharyngeal pressure generated by nHF is (i) similar to, or below, that generated during CPAP and (ii) increases with increased flow rate.<sup>13,16–18</sup> Spence *et al.* found a linear relationship between flow and intrapharyngeal pressure,<sup>18</sup> and Wilkinson *et al.* found that pharyngeal pressure increased by 0.8 cm H<sub>2</sub>O per 1 L/min increase in flow.<sup>13</sup> Others

have demonstrated similar trends but with smaller pressure changes with increasing flow.<sup>17</sup>

For a given flow, the delivered pressure appears to be inversely proportional to an infant's weight,<sup>13,16</sup> that is, the smallest infants require less flow to generate the same distending pressure as larger infants. One small study compared pressure generated by two nHF devices, Optiflow (Fisher & Paykel, Auckland, New Zealand) and Vapotherm (Vapotherm, Exeter, CA, USA), and showed no pressure differences with flows of 2–6 L/min and slightly higher pharyngeal pressures with Vapotherm at 8 L/min.<sup>14</sup> This may reflect the inclusion of a pressure relief valve in the Optiflow system but not in Vapotherm.<sup>14</sup>

Distending pressure is also determined by leak;<sup>11</sup> manufacturers recommend a prong size that allows 50% 'leak' around the nares. Mouth leaks also potentially affect pressure generation, although findings from studies investigating this are contradictory.<sup>13,16</sup> Distending pressure increases as the leak around the prongs falls; hence, appropriate sizing of prongs is an important safety consideration for nHF use.

### Nasopharyngeal dead space washout

Washout of the nasopharyngeal dead space and subsequent carbon dioxide removal have been proposed as a mechanism of action of nHF,<sup>15</sup> although no neonatal studies have been performed to confirm this.<sup>19,20</sup> A study of neonatal piglets showed that, during nHF, carbon dioxide levels reduced in a flow-dependent manner, independent of leak.<sup>19</sup> A recent bench-top study with a simulated preterm infant lung model reported faster gas washout during nHF than during CPAP over a variety of flow and pressure settings,<sup>20</sup> potentially due to more turbulent gas flow during nHF.

Studies in children with bronchiolitis have shown similar reductions in carbon dioxide levels after nHF commencement;<sup>21,22</sup> however, studies in paediatric patients post-cardiac surgery have not shown this effect.<sup>23</sup>

### Heating and humidification of gases

Provision of heated and humidified gas flow is a requirement for all respiratory support modes. Delivery of unheated, unhumidified gas can adversely affect the airway mucosa and mucociliary function, reducing pulmonary compliance and inducing bronchoconstriction.<sup>24,25</sup> All commercially available nHF systems therefore deliver heated, humidified gas.<sup>21</sup> A laboratory study comparing the temperature and humidity achieved during nHF and CPAP, using a simulated neonatal lung model, showed that higher mean gas temperatures were achieved during CPAP (34.5 vs. 34.0°C,  $P < 0.01$ ), but higher relative humidity was achieved with nHF (83 vs. 76%,  $P < 0.01$ ).<sup>26</sup> Relative humidity decreased slightly as flow increased during CPAP, whereas it increased with increasing flow during nHF.<sup>26</sup> A neonatal manikin study<sup>27</sup> also reported that overall temperature was higher with CPAP than nHF (35.5 vs. 34.3°C) but conversely showed that relative humidity decreased during nHF at the highest flow levels. All the devices produced absolute humidity within international recommendations at flows of 8 L/min. The clinical significance of minor differences between devices is uncertain.

### Reduction of work of breathing

The large surface area of the nasopharynx allows humidification and heating of inspired gas but causes resistance to inspiratory flow.<sup>15</sup> CPAP reduces this resistance by splinting the upper airway open.<sup>28</sup> nHF delivers gas flows above the peak inspiratory flow of the patient, reducing resistance and work of breathing (WOB).<sup>15</sup> Some studies that have compared the effects of nHF and CPAP on WOB in neonates have found that both modes reduce respiratory effort,<sup>29–31</sup> although de Jongh reported lower indices of WOB with CPAP 5–6 cm H<sub>2</sub>O compared with nHF of 3–5 L/min.<sup>32</sup>

## Indications for nHF Therapy

### Primary respiratory support

In 2016, a Cochrane review pooled the available data comparing nHF with CPAP for initial treatment of respiratory distress syndrome (RDS) in preterm infants (four trials, 439 patients).<sup>12,33–36</sup> With the exception of one small study,<sup>36</sup> infants were all >28 weeks' gestation. One study<sup>35</sup> permitted surfactant treatment using the INSURE technique (intubation-surfactant-extubation) before and after study entry. The review concluded that rates of treatment failure and need for intubation within 7 days were similar between nHF and CPAP, with no difference in rates of death or BPD.

Since 2016, three further non-inferiority trials have added to the evidence base. While one trial found nHF to be non-inferior to CPAP, the largest favoured nCPAP, and a third, smaller study was inconclusive. Lavizzari *et al.* found nHF to be non-inferior to CPAP with respect to the need for intubation and ventilation in 316 infants between 29 and 36 + 6 weeks' gestation,<sup>37</sup> with no differences in secondary outcomes. Treatment failure occurred in 10.8 versus 9.5% of infants receiving nHF and CPAP, respectively (95% confidence interval (CI) of risk difference, –6.0, 8.6,  $P = 0.71$ ). The multicentre High Flow Nasal Cannulae as Primary Support in the Treatment of Early Respiratory Distress (HIPSTER) trial conducted by Roberts *et al.* randomised 564 infants at 28–36 + 6 weeks' gestation to nHF or CPAP.<sup>38</sup> The trial was stopped after 75% of the planned recruitment due to a significantly higher treatment failure rate (defined by objective oxygenation, blood gas and apnoea criteria) in the nHF group (25.5 vs. 13.3% in the CPAP group,  $P < 0.001$ ). Intubation rates were similar between groups (15.5 nHF vs. 11.5% CPAP,  $P = 0.17$ ), including in the <32 week gestational age (GA) subgroup. Infants in the nHF group had less nasal trauma (18.5 vs. 8.3%,  $P < 0.001$ ) and less pneumothoraces while on allocated treatment (2.1 vs. 0.0%,  $P = 0.02$ ); however, the median duration of respiratory support was a day longer in the nHF group. A third, smaller randomised controlled trial (RCT) by Shin *et al.*<sup>39</sup> randomised 85 infants (mean 33 weeks' GA) to nHF or CPAP and was technically inconclusive, although the rate of treatment failure was 38% in the nHF group versus 21% in the CPAP group (95% CI of risk difference –1.9, 36.2,  $P = 0.1$ ). There were no significant differences in secondary clinical outcomes.

There were important differences in the design of these three studies; the use of surfactant and 'rescue' CPAP in the nHF arms varied. In the Lavizzari trial, infants could receive surfactant before or after trial enrolment, without fulfilling treatment failure criteria, whereas infants in the other trials could not. Almost half

of the infants included in Lavizzari's trial received surfactant (nHF: 44%, CPAP 46%). Rates of surfactant administration were similar in both groups; however, its use may in part explain the lower rate of treatment failure compared with the other trials, where no infants received surfactant prior to reaching treatment failure criteria. A further difference between the studies is that infants in the Lavizzari study who met failure criteria were intubated and ventilated, whereas those in the nHF arm of the HIPSTER trial were given 'rescue' CPAP. This prevented intubation in 40% of those who reached treatment failure in the nHF group. Use of 'rescue' CPAP in the nHF group may have influenced the secondary outcomes in this trial and is an important consideration for units instituting nHF without having CPAP available as a rescue therapy. Finally, the study by Shin *et al.* was much smaller and had a much larger non-inferiority margin than the others.

Differences between these studies make a universal recommendation regarding the use of nHF for primary respiratory support difficult. A 2015 meeting of international experts concluded that, while nHF can be considered an alternative to CPAP, further large RCTs are required to determine its efficacy and safety as primary support for RDS.<sup>40</sup> Recently, a panel of international research experts was surveyed to determine consensus guidelines for the management of neonates with nHF.<sup>41</sup> There was general support for the use of nHF as primary therapy for respiratory distress, with consideration of the infant's gestational age and initial oxygen requirement. Individual units may consider these and other factors, such as rate of surfactant use, patient population and local rate of nasal trauma, when choosing between primary support therapies.

In summary, current evidence for the use of nHF for primary respiratory support of preterm RDS suggests the following:

- 1 nHF may not be equivalent to CPAP.
- 2 The use of 'rescue' CPAP in the event of nHF failure avoids intubation in many infants.
- 3 There is little evidence to guide primary nHF therapy in infants less than 28 weeks' GA.

### Post-extubation support

Most very preterm infants require respiratory support post-extubation. The 2016 Cochrane meta-analysis of nHF post-extubation (six trials, 934 infants) concluded that there was no difference between nHF and CPAP with respect to treatment failure, reintubation or secondary outcomes, including BPD, in preterm infants >28 weeks' gestation.<sup>12</sup> Infants randomised to nHF had less nasal trauma and a borderline lower pneumothorax rate. Some studies allowed 'rescue' CPAP prior to reintubation<sup>33,42–45</sup> or crossover from CPAP to nHF in the event of nasal trauma.<sup>42,43</sup> The largest study of preterm infants reported that 48% of infants who reached treatment failure with nHF were successfully treated with 'rescue' CPAP, avoiding reintubation for at least 7 days.<sup>44</sup>

Two more recent RCTs have similarly found no difference in treatment failure or reintubation rates between nHF and CPAP post-extubation. Kang *et al.* randomised 161 preterm infants <32 weeks' GA<sup>42</sup> and found no difference in these outcomes overall, but treatment failure was higher with nHF in the smaller infants (26–28<sup>+</sup> weeks' GA). A second, smaller study found a higher rate of treatment failure but a similar rate of intubation in

infants <32 weeks' GA managed with nHF;<sup>46</sup> the intubation rates were similar due to the use of rescue CPAP in the nHF arm.

Similar to the primary support trials, relatively few infants <28 weeks' GA have been studied (233 infants from two studies included in the Cochrane meta-analysis). The largest trial that included infants <28 weeks' GA reported a much higher treatment failure rate for nHF in infants <26 weeks' GA, and the authors advised caution in its use as post-extubation support in these infants.<sup>44</sup>

Several retrospective cohort studies have demonstrated a longer duration of respiratory support or supplemental oxygen therapy in very preterm infants managed with nHF compared with CPAP.<sup>47,48</sup> The interpretation of these studies is limited by methodological considerations; there may be significant differences in the severity of lung disease and other comorbidities between the groups. Some RCTs have also shown longer duration of respiratory support with nHF use;<sup>33,38</sup> yet others have not.<sup>37,39</sup> Importantly, there is no corresponding increase in the incidence of BPD for infants managed with nHF.<sup>12</sup> Further research may inform the best approach to weaning nHF to minimise the duration of respiratory support for preterm infants.

In summary:

- 1 nHF has similar efficacy to CPAP for post-extubation support with respect to reintubation, death and BPD, with the exception of babies <28 weeks' gestation, in whom CPAP should be used by preference
- 2 Following extubation, nHF is associated with less nasal trauma than CPAP.
- 3 Availability of rescue CPAP may be important for preventing extubation failure when using nHF and may be preferred as first-line therapy in the smallest infants.

### Weaning from CPAP

Four RCTs have examined nHF as a strategy for weaning from CPAP, with conflicting effects seen on the duration of supplemental oxygen and respiratory support.<sup>49–52</sup> Two studies<sup>51,52</sup> found no difference in outcome when weaning directly from CPAP was compared with changing to nHF. The third study<sup>49</sup> found that using nHF to wean from CPAP resulted in more days of supplemental oxygen and respiratory support, and a fourth study found the reverse to be true.<sup>50</sup> Currently, there is insufficient evidence to guide practice recommendations regarding the use of nHF for weaning from CPAP.

### Use in non-tertiary centres

Non-tertiary centres have unique considerations when choosing between non-invasive respiratory support modes, such as nursing skills, availability of rescue CPAP and potential need to transfer deteriorating babies. Thus far, all the evidence for neonatal nHF use comes from tertiary neonatal units. However, a large, multi-centre RCT comparing nHF with CPAP in Australian non-tertiary neonatal units has recently completed recruitment,<sup>53</sup> and results are awaited. While cost is a consideration, there is little evidence that nHF is cheaper than CPAP. In the HIPSTER trial, total cost of tertiary hospital stay did not differ significantly between groups,<sup>38</sup> and a recent economic evaluation of nHF found that the higher

capital equipment costs of CPAP were not outweighed by the higher consumable costs of nHF.<sup>54</sup>

## Future Research Directions

There are a number of potential novel applications for neonatal nHF therapy that are worthy of evaluation in a research setting:

### Intubation support

Neonatal intubation is a difficult skill to perform; attempts commonly take longer than recommended<sup>55</sup> and result in desaturation or bradycardia.<sup>56</sup> A new technique called Transnasal Humidified Rapid-Insufflation Ventilatory Exchange utilises nHF to prolong the safe apnoeic time following induction of anaesthesia,<sup>57</sup> likely due to upper airway splinting, continuous insufflation and dead space washout.<sup>57</sup> This technique has recently been evaluated in healthy children undergoing anaesthesia<sup>58</sup> and was found to significantly prolong the time to desaturation. No studies have investigated this technique in preterm infants who frequently become unstable during intubation.

### nHF for delivery room stabilisation of preterm infants

nHF may be of value as a method of stabilisation for preterm infants in the delivery room. It is quick and easy to apply, and it is postulated that it may stimulate respiration and improve gas exchange,<sup>59</sup> avoiding the need for immediate CPAP or positive pressure ventilation. nHF use in the delivery room has only been reported in a single, small, observational study of infants <30 weeks' GA<sup>59</sup>; most infants were successfully stabilised with nHF, and overall intubation rates were lower than the unit average prior to the study. The use of delivery room nHF has yet to be evaluated in adequately powered randomised trials.

### nHF during neonatal transport

The use of nHF in elective and emergency neonatal transport has been reported<sup>60</sup> in one UK study of 102 infants. Whilst no comparison was made with transports on CPAP, the authors reported a very small (2%) increase in oxygen requirement during transport on nHF, and it was believed to be safe and effective. Provision of 'rescue' CPAP during transport in the event of patient deterioration may present a logistic challenge; therefore, this requires further exploration before widespread implementation.

Other avenues for future research into nHF use in preterm infants include:

- Comparisons between commercially available nHF devices.
- Investigating the best approach to weaning nHF, particularly given concerns about the duration of supplemental oxygen and respiratory support.
- Further evaluation of nHF use in extremely preterm infants.

## Conclusions

nHF therapy is a relatively new addition to the neonatal intensive care environment. There is good evidence for its use as an

alternative to CPAP for post-extubation respiratory support; however, CPAP should be available as 'rescue' therapy in the event of nHF failure. Caution should also be applied when using nHF in extremely preterm infants. Evidence for primary respiratory support and weaning from CPAP is less clear-cut. As nHF use in tertiary and non-tertiary centres increases, further research is required regarding specific aspects of the therapy and to explore new therapeutic indications.

## Multiple Choice Questions

1 Current evidence supports the efficacy of nasal high flow (nHF) compared with continuous positive airway pressure (CPAP) for which of the following indications in preterm neonates?

- Primary respiratory support
- Post-extubation support
- During intubation
- 'Weaning' from CPAP
- Delivery room stabilisation

Answer: b. Current meta-analyses indicate that the use of nHF for post-extubation support of preterm infants results in similar rates of reintubation compared with CPAP. The exception to this is infants <26 weeks' gestation, for whom the limited data suggest that rates of treatment failure with nHF are high. Studies have shown conflicting results for nHF use as primary respiratory support and 'weaning' from CPAP. The use of nHF during intubation and for delivery room stabilisation requires further investigation.

2 Regarding the use of nHF for neonates:

- Similar distending pressure is generated irrespective of the infant's weight.
- nHF interfaces use larger nasal prongs than CPAP.
- Distending pressure decreases with increasing flow rates.
- Prongs should be sized to completely occlude the nostrils.
- nHF may increase nasopharyngeal dead space washout.

Answer: e. Nasopharyngeal dead space washout is a proposed mechanism of action of nHF. Other mechanisms of action include generation of continuous distending pressure, which increases with increasing flow rate and is inversely proportional to infant weight. To avoid excessive distending pressure, nHF prongs (which are smaller than CPAP prongs) should be sized to allow some gas leak around them in the nares.

3 What advantage does the use of nHF offer compared with CPAP?

- nHF is cheaper.
- nHF reduces nasal trauma.
- nHF reduces the risk of chronic lung disease.
- nHF devices allow clinicians to set a specific distending pressure.
- nHF does not require heated humidified gas.

Answer: b. nHF reduces the rates of nasal trauma compared with CPAP. Current evidence demonstrates no difference in cost between nHF and CPAP when all factors are considered. nHF devices (which all use heated humidified gas) require flow, not pressure, to be set. Current evidence indicates that the rate of

chronic lung disease is similar when nHF is used compared with CPAP.

## References

- Fujiwara T, Maeta H, Chida S, Morita T, Watabe Y, Abe T. Artificial surfactant therapy in hyaline-membrane disease. *Lancet* 1980; **1**: 55–9.
- Avery ME, Mead J. Surface properties in relation to atelectasis and hyaline membrane disease. *AMA Am. J. Dis. Child.* 1959; **97** (5 Pt 1): 517–23.
- Merritt TA, Hallman M, Bloom BT et al. Prophylactic treatment of very premature infants with human surfactant. *N. Engl. J. Med.* 1986; **315**: 785–90.
- Avery ME, Tooley WH, Keller JB et al. Is chronic lung disease in low birth weight infants preventable? A survey of eight centers. *Pediatrics* 1987; **79**: 26–30.
- Van Marter LJ, Allred EN, Pagano M et al. Do clinical markers of barotrauma and oxygen toxicity explain interhospital variation in rates of chronic lung disease? The Neonatology Committee for the Developmental Network. *Pediatrics* 2000; **105**: 1194–201.
- Dunn MS, Kaempf J, de Klerk A et al. Randomized trial comparing 3 approaches to the initial respiratory management of preterm neonates. *Pediatrics* 2011; **128**: e1069–76.
- Finer NN, Carlo WA, Walsh MC et al. Early CPAP versus surfactant in extremely preterm infants. *N. Engl. J. Med.* 2010; **362**: 1970–9.
- Morley CJ, Davis PG, Doyle LW, Brion LP, Hascoet JM, Carlin JB. Nasal CPAP or intubation at birth for very preterm infants. *N. Engl. J. Med.* 2008; **358**: 700–8.
- Fischer HS, Buhner C. Avoiding endotracheal ventilation to prevent bronchopulmonary dysplasia: A meta-analysis. *Pediatrics* 2013; **132**: e1351–60.
- Schmolzer GM, Kumar M, Pichler G, Aziz K, O'Reilly M, Cheung PY. Non-invasive versus invasive respiratory support in preterm infants at birth: Systematic review and meta-analysis. *BMJ* 2013; **347**: f5980.
- Imbulana DI, Manley BJ, Dawson JA, Davis PG, Owen LS. Nasal injury in preterm infants receiving non-invasive respiratory support: A systematic review. *Arch. Dis. Child. Fetal Neonatal Ed.* 2018; **103**: F29–35.
- Wilkinson D, Andersen C, O'Donnell CP, De Paoli AG, Manley BJ. High flow nasal cannula for respiratory support in preterm infants. *Cochrane Database Syst. Rev.* 2016; **2**: CD006405.
- Wilkinson DJ, Andersen CC, Smith K, Holberton J. Pharyngeal pressure with high-flow nasal cannulae in premature infants. *J. Perinatol.* 2008; **28**: 42–7.
- Collins CL, Holberton JR, Konig K. Comparison of the pharyngeal pressure provided by two heated, humidified high-flow nasal cannulae devices in premature infants. *J. Paediatr. Child Health* 2013; **49**: 554–6.
- Dysart K, Miller TL, Wolfson MR, Shaffer TH. Research in high flow therapy: Mechanisms of action. *Respir. Med.* 2009; **103**: 1400–5.
- Kubiccka ZJ, Limauro J, Darnall RA. Heated, humidified high-flow nasal cannula therapy: Yet another way to deliver continuous positive airway pressure? *Pediatrics* 2008; **121**: 82–8.
- Lampland AL, Plumm B, Meyers PA, Worwa CT, Mammel MC. Observational study of humidified high-flow nasal cannula compared with nasal continuous positive airway pressure. *J. Pediatr.* 2009; **154**: 177–82.
- Spence KL, Murphy D, Kilian C, McGonigle R, Kilani RA. High-flow nasal cannula as a device to provide continuous positive airway pressure in infants. *J. Perinatol.* 2007; **27**: 772–5.
- Frizzola M, Miller TL, Rodriguez ME et al. High-flow nasal cannula: Impact on oxygenation and ventilation in an acute lung injury model. *Pediatr. Pulmonol.* 2011; **46**: 67–74.
- Sivieri EM, Foglia EE, Abbasi S. Carbon dioxide washout during high flow nasal cannula versus nasal CPAP support: An in vitro study. *Pediatr. Pulmonol.* 2017; **52**: 792–8.
- Bressan S, Balzani M, Krauss B, Pettenazzo A, Zanonato S, Baraldi E. High-flow nasal cannula oxygen for bronchiolitis in a pediatric ward: A pilot study. *Eur. J. Pediatr.* 2013; **172**: 1649–56.
- Kallappa C, Hufton M, Millen G, Ninan TK. Use of high flow nasal cannula oxygen (HFNCO) in infants with bronchiolitis on a paediatric ward: A 3-year experience. *Arch. Dis. Child.* 2014; **99**: 790–1.
- Testa G, Iodice F, Ricci Z et al. Comparative evaluation of high-flow nasal cannula and conventional oxygen therapy in paediatric cardiac surgical patients: A randomized controlled trial. *Interact. Cardiovasc. Thorac. Surg.* 2014; **19**: 456–61.
- Fontanari P, Burnet H, Zattara-Hartmann MC, Jammes Y. Changes in airway resistance induced by nasal inhalation of cold dry, dry, or moist air in normal individuals. *J. Appl. Physiol.* 1996; **81**: 1739–43.
- Sottiaux TM. Consequences of under- and over-humidification. *Respir. Care Clin. N. Am.* 2006; **12**: 233–52.
- Chang GY, Cox CA, Shaffer TH. Nasal cannula, CPAP, and high-flow nasal cannula: Effect of flow on temperature, humidity, pressure, and resistance. *Biomed. Instrum. Technol.* 2011; **45**: 69–74.
- Roberts CT, Kortekaas R, Dawson JA, Manley BJ, Owen LS, Davis PG. The effects of non-invasive respiratory support on oropharyngeal temperature and humidity: A neonatal manikin study. *Arch. Dis. Child. Fetal Neonatal Ed.* 2016; **101**: F248–52.
- Chowdhury O, Wedderburn CJ, Duffy D, Greenough A. CPAP review. *Eur. J. Pediatr.* 2012; **171**: 1441–8.
- Saslow JG, Aghai ZH, Nakhla TA et al. Work of breathing using high-flow nasal cannula in preterm infants. *J. Perinatol.* 2006; **26**: 476–80.
- Shetty S, Hickey A, Rafferty GF, Peacock JL, Greenough A. Work of breathing during CPAP and heated humidified high-flow nasal cannula. *Arch. Dis. Child. Fetal Neonatal Ed.* 2016; **101**: F404–7.
- Lavizzari A, Veneroni C, Colnaghi M et al. Respiratory mechanics during NCPAP and HHHFNC at equal distending pressures. *Arch. Dis. Child. Fetal Neonatal Ed.* 2014; **99**: F315–20.
- de Jongh BE, Locke R, Mackley A et al. Work of breathing indices in infants with respiratory insufficiency receiving high-flow nasal cannula and nasal continuous positive airway pressure. *J. Perinatol.* 2014; **34**: 27–32.
- Yoder BA, Stoddard RA, Li M, King J, Dimberger DR, Abbasi S. Heated, humidified high-flow nasal cannula versus nasal CPAP for respiratory support in neonates. *Pediatrics* 2013; **131**: e1482–90.
- Ciuffini F, Pietrasanta C, Lavizzari A et al. Comparison between two different modes of non-invasive ventilatory support in preterm newborn infants with respiratory distress syndrome mild to moderate: Preliminary data. *Pediatr. Med. Chir.* 2014; **36**: 88.
- Iranpour R, Sadeghnia A, Hesarak M. High-flow nasal cannula versus nasal continuous positive airway pressure in the management of respiratory distress syndrome. *J. Isfahan Med. Sch.* 2011; **29**: 1.
- Nair G, Karna P. Comparison of the effects of Vapotherm and nasal CPAP in respiratory distress. *Pediatr. Acad. Soc.* 2005; **57**: 2054.
- Lavizzari A, Colnaghi M, Ciuffini F et al. Heated, humidified high-flow nasal cannula vs nasal continuous positive airway pressure for respiratory distress syndrome of prematurity: A randomized clinical noninferiority trial. *JAMA Pediatr.* 2016; <https://doi.org/10.1001/jamapediatrics.2016.1243> (forthcoming).
- Roberts CT, Owen LS, Manley BJ et al. Nasal high-flow therapy for primary respiratory support in preterm infants. *N. Engl. J. Med.* 2016; **375**: 1142–51.
- Shin J, Park K, Lee EH, Choi BM. Humidified high flow nasal cannula versus CPAP as initial respiratory support in preterm infants. *J. Korean Med. Sci.* 2017; **32**: 650–5.
- Roehr CC, Yoder BA, Davis PG, Ives K. Evidence support and guidelines for using heated, humidified, high-flow nasal cannulae in neonatology: Oxford nasal high-flow therapy meeting, 2015. *Clin. Perinatol.* 2016; **43**: 693–705.

- 41 Yoder BA, Manley B, Collins C *et al.* Consensus approach to nasal high-flow therapy in neonates. *J. Perinatol.* 2017; **37**: 809–13.
- 42 Kang WQ, Xu BL, Liu DP *et al.* Efficacy of heated humidified high-flow nasal cannula in preterm infants aged less than 32 weeks after ventilator weaning. *Zhongguo Dang Dai Er Ke Za Zhi* 2016; **18**: 488–91.
- 43 Collins CL, Holberton JR, Barfield C, Davis PG. A randomized controlled trial to compare heated humidified high-flow nasal cannulae with nasal continuous positive airway pressure postextubation in premature infants. *J. Pediatr.* 2013; **162**: 949–54. e1.
- 44 Manley BJ, Owen LS, Doyle LW *et al.* High-flow nasal cannulae in very preterm infants after extubation. *N. Engl. J. Med.* 2013; **369**: 1425–33.
- 45 Mostafa-Gharehbaghi M, Mojabi H. Comparing the effectiveness of nasal continuous positive airway pressure (NCPAP) and high flow nasal cannula (HFNC) in prevention of post extubation assisted ventilation. *Zahedan J. Res. Med. Sci.* 2015; **17**: e984.
- 46 Soonsawad S, Swatesutipun B, Limrungsikul A, Nuntnarumit P. Heated humidified high-flow nasal cannula for prevention of extubation failure in preterm infants. *Indian J. Pediatr.* 2017; **84**: 262–6.
- 47 Hoffman SB, Terrell N, Driscoll CH, Davis NL. Impact of high-flow nasal cannula use on neonatal respiratory support patterns and length of stay. *Respir. Care* 2016; **61**: 1299–304.
- 48 Taha DK, Kornhauser M, Greenspan JS, Dysart KC, Aghai ZH. High flow nasal cannula use is associated with increased morbidity and length of hospitalization in extremely low birth weight infants. *J. Pediatr.* 2016; **173**: 50–5. e1.
- 49 Abdel-Hady H, Shouman B, Aly H. Early weaning from CPAP to high flow nasal cannula in preterm infants is associated with prolonged oxygen requirement. *Early Hum. Dev.* 2011; **87**: 205–8.
- 50 Badiie Z, Eshghi A, Mohammadzadeh M. High flow nasal cannula as a method for rapid weaning from nasal continuous positive airway pressure. *Int. J. Prev. Med.* 2015; **6**: 33.
- 51 Tang J, Reid S, Lutz T, Malcolm G, Oliver S, Osborn DA. Randomised controlled trial of weaning strategies for preterm infants on nasal continuous positive airway pressure. *BMC Pediatr.* 2015; **15**: 147.
- 52 Soonsawad S, Tongsaawang N, Nuntnarumit P. Heated humidified high-flow nasal cannula for weaning from continuous positive airway pressure in preterm infants: A randomized controlled trial. *Neonatology* 2016; **110**: 204–9.
- 53 Manley BJ, Roberts CT, Arnolda GRB *et al.* A multicentre, randomised controlled, non-inferiority trial, comparing nasal high flow with nasal continuous positive airway pressure as primary support for newborn infants with early respiratory distress born in Australian non-tertiary special care nurseries (the HUNTER trial): Study protocol. *BMJ Open* 2017; **7**: e016746.
- 54 Fleeman N, Mahon J, Bates V *et al.* The clinical effectiveness and cost-effectiveness of heated humidified high-flow nasal cannula compared with usual care for preterm infants. *Health Technol. Assess.* 2016; **20**: 1–68.
- 55 O'Donnell CP, Kamlin CO, Davis PG, Morley CJ. Endotracheal intubation attempts during neonatal resuscitation: Success rates, duration, and adverse effects. *Pediatrics* 2006; **117**: e16–21.
- 56 Hatch LD, Grubb PH, Lea AS *et al.* Endotracheal intubation in neonates: A prospective study of adverse safety events in 162 infants. *J. Pediatr.* 2016; **168**: 62–6. e6.
- 57 Patel A, Nouraei SA. Transnasal humidified rapid-insufflation ventilatory exchange (THRIVE): A physiological method of increasing apnoea time in patients with difficult airways. *Anaesthesia* 2015; **70**: 323–9.
- 58 Humphreys S, Lee-Archer P, Reyne G, Long D, Williams T, Schibler A. Transnasal humidified rapid-insufflation ventilatory exchange (THRIVE) in children: A randomized controlled trial. *Br. J. Anaesth.* 2017; **118**: 232–8.
- 59 Reynolds P, Leontiadi S, Lawson T, Otunla T, Ejiwumi O, Holland N. Stabilisation of premature infants in the delivery room with nasal high flow. *Arch. Dis. Child. Fetal Neonatal Ed.* 2016; **101**: F284–7.
- 60 Boyle MA, Dhar A, Chaudhary R *et al.* Introducing high-flow nasal cannula to the neonatal transport environment. *Acta Paediatr.* 2017; **106**: 509–12.